

Certificado No. (Certificate #): 53119 Equipo (Instrument): Reloj Indicador
Fecha de Calibración (Calibration Date): 2025-01-21 Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.0500 in	0.0502 in	0.0002 in	± 0.0003 in	± 6.8E-05 in
0.1005 in	0.1007 in	0.0002 in	± 0.0003 in	± 9.1E-05 in
0.1050 in	0.1052 in	0.0002 in	± 0.0003 in	± 9.4E-05 in
0.1100 in	0.1102 in	0.0002 in	± 0.0003 in	± 9.6E-05 in
0.1500 in	0.1502 in	0.0002 in	± 0.0003 in	± 1.7E-04 in
0.2000 in	0.2002 in	0.0002 in	± 0.0003 in	± 1.9E-04 in
0.3000 in	0.3002 in	0.0002 in	± 0.0003 in	± 2.2E-04 in
0.4000 in	0.4002 in	0.0002 in	± 0.0003 in	± 2.9E-04 in
0.5000 in	0.5001 in	0.0001 in	± 0.0003 in	± 3.5E-04 in