

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
130.3 μm	130.8 μm	0.5 μm	± 5 μm	±	5.8E-01 μm
523.0 μm	523.6 μm	0.6 μm	± 5 μm	±	5.8E-01 μm