

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.975 °C	31.1 °C	0.125 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.517 %HR	32 %HR	2.483 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.816 %HR	52 %HR	1.184 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	72.012 %HR	75 %HR	2.988 %HR	± 5.0 %HR	± 1.9E+00 %HR