

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.975 °C	30.8 °C	-0.175 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.517 %HR	32 %HR	2.483 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.816 %HR	49 %HR	-1.816 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	72.012 %HR	71 %HR	-1.012 %HR	± 5.0 %HR	± 1.9E+00 %HR