

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.5	0.5000	0.780	± 1.6	± 2.9E+00 mm					
0 a 1.01	1.0000	10.750	± 2	± 2.9E+00 mm					
0 a 1.51	1.5000	10.730	± 2.4	± 2.9E+00 mm					
0 a 2.01	2.0000	10.710	± 2.8	± 2.9E+00 mm					
0 a 4.01	4.0000	10.620	± 4.4	± 2.9E+00 mm					
0 a 6.01	6.0000	10.530	± 6	± 2.9E+00 mm					
0 a 8.01	8.0000	10.440	± 7.6	± 2.9E+00 mm					
0 a 10.01	10.0000	10.350	± 9.2	± 2.9E+00 mm					
0 a 15.01	15.0000	10.140	± 13.2	± 2.9E+00 mm					
0 a 25.01	25.0000	9.750	± 21.2	± 2.9E+00 mm					
0 a 30.01	30.0000	9.570	± 25.2	± 2.9E+00 mm					
0 a 40.02	40.0000	19.230	± 33.2	± 2.9E+00 mm					
0 a 50.02	50.0000	18.930	± 41.2	± 2.9E+00 mm					
0 a 60.02	60.0000	20.000	± 49.2	± 3.0E+00 mm					
0 a 70.02	70.0000	20.000	± 57.2	± 3.0E+00 mm					
0 a 80.02	80.0000	20.000	± 65.2	± 2.9E+00 mm					
0 a 90.02	90.0000	20.000	± 73.2	± 2.9E+00 mm					
0 a 100.02	100.0000	20.000	± 81.2	± 2.9E+00 mm					