

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.935 °C	30.9 °C	-0.035 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.770 %HR	31 %HR	2.230 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.184 %HR	51 %HR	-0.184 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.501 %HR	71 %HR	-0.501 %HR	± 5.0 %HR	± 1.9E+00 %HR