

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.935 °C	30.4 °C	-0.535 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.770 %HR	32 %HR	3.230 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.184 %HR	52 %HR	0.816 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	71.501 %HR	76 %HR	4.499 %HR	± 5.0 %HR	± 1.9E+00 %HR