

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.313 °C	30.3 °C	-0.013 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.0 %HR	0.737 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	50.0 %HR	-0.754 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 2.0 %HR	± 1.9E+00 %HR