

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.098	2.25	± 1.2	± 4.3E-01 mm					
0 a 0.15	0.148	2.25	± 1.3	± 4.3E-01 mm					
0 a 0.2	0.198	2.25	± 1.3	± 4.3E-01 mm					
0 a 0.3	0.298	2.26	± 1.4	± 4.3E-01 mm					
0 a 0.4	0.398	2.26	± 1.5	± 4.3E-01 mm					
0 a 0.5	0.498	2.26	± 1.6	± 4.3E-01 mm					
0 a 0.6	0.598	2.27	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.698	2.27	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.798	2.27	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.898	2.28	± 1.9	± 4.3E-01 mm					
0 a 1	0.998	2.28	± 2	± 4.3E-01 mm					
0 a 1.1	1.098	2.28	± 2	± 4.8E-01 mm					
0 a 1.2	1.198	2.29	± 2.1	± 4.8E-01 mm					