

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
18 °C	18.130 °C	17.6 °C	-0.530 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	19.5 °C	-0.665 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.7 °C	-0.349 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.358 °C	27.8 °C	-0.558 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.795 %HR	46.7 %HR	5.905 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	52.9 %HR	2.146 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.4 %HR	-3.048 %HR	± 3.0 %HR	± 1.9E+00 %HR