

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
18 °C	18.130 °C	17.9 °C	-0.230 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	19.9 °C	-0.265 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.6 °C	-0.449 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.358 °C	28.6 °C	0.242 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.795 %HR	47.8 %HR	7.005 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	57.3 %HR	6.546 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.223 %HR	61.1 %HR	0.877 %HR	± 3.0 %HR	± 1.9E+00 %HR