

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
18 °C	18.130 °C	17.6 °C	-0.530 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	19.5 °C	-0.665 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.5 °C	-0.549 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.358 °C	28.6 °C	0.242 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.795 %HR	47.8 %HR	7.005 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	57.3 %HR	6.546 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.3 %HR	0.852 %HR	± 3.0 %HR	± 1.9E+00 %HR