

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
21 °C	21.138 °C	21.6 °C	0.462 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.143 °C	23.5 °C	0.357 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.6 °C	-0.449 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.358 °C	28.6 °C	0.242 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.795 %HR	47.8 %HR	7.005 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	57.3 %HR	6.546 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.223 %HR	61.1 %HR	0.877 %HR	± 3.0 %HR	± 1.9E+00 %HR