

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
21 °C	21.138 °C	20.7 °C	-0.438 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.143 °C	22.8 °C	-0.343 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.8 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.358 °C	30.1 °C	1.742 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.795 %HR	37.0 %HR	-3.795 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.754 %HR	52.8 %HR	2.046 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	60.223 %HR	62.7 %HR	2.477 %HR	± 3.0 %HR	± 1.9E+00 %HR