

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.05035	-0.350	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.10036	-0.360	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.15031	-0.310	± 0.6	± 4.3E-01 mm					
0 a 1	0.99962	0.380	± 1	± 4.3E-01 mm					
0 a 2	2.00011	-0.110	± 1.4	± 4.8E-01 mm					
0 a 3	2.99998	0.020	± 1.8	± 4.8E-01 mm					
0 a 4	4.00034	-0.340	± 2.2	± 4.8E-01 mm					
0 a 5	4.99985	0.150	± 2.6	± 4.8E-01 mm					
0 a 6	5.99978	0.220	± 3	± 3.6E-01 mm					
0 a 7	6.99959	0.410	± 3.4	± 3.6E-01 mm					
0 a 8	7.99963	0.370	± 3.8	± 3.6E-01 mm					
0 a 9	8.99974	0.260	± 4.2	± 3.6E-01 mm					
0 a 10	9.99998	0.020	± 4.6	± 3.6E-01 mm					
0 a 11	10.99956	0.440	± 5	± 3.6E-01 mm					
0 a 12	11.99983	0.170	± 5.4	± 3.6E-01 mm					
0 a 13	12.99989	0.110	± 5.8	± 3.6E-01 mm					
0 a 14	14.00028	-0.280	± 6.2	± 3.6E-01 mm					
0 a 15	15.00016	-0.160	± 6.6	± 3.6E-01 mm					