

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.201 °C                  | 15.2 °C            | -0.011 °C              | ± 0.1 °C                      | ± 4.0E-01 °C                         |
| 20 °C                      | 20.035 °C                  | 20.0 °C            | -0.035 °C              | ± 0.1 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.363 °C                  | 30.5 °C            | 0.087 °C               | ± 0.1 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.570 %HR                 | 31.6 %HR           | 1.030 %HR              | ± 1.5 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.212 %HR                 | 51.2 %HR           | 0.988 %HR              | ± 1.5 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.857 %HR                 | 70.8 %HR           | -0.057 %HR             | ± 1.5 %HR                     | ± 1.9E+00 %HR                        |