

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.165 °C	20.2 °C	0.035 °C	± 0.3 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	34.8 %HR	-0.371 %HR	± 1.5 %HR	± 1.9E+00 %HR
50 %HR	50.734 %HR	51.3 %HR	0.566 %HR	± 1.5 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	75.0 %HR	-0.560 %HR	± 1.5 %HR	± 1.9E+00 %HR