

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) | |
|--------------------------|---------------------|----------------------|--------------------------|------------------------------------|-----------------------|
| 130.3 μm | 129.8 μm | -0.5 μm | $\pm 5 \mu\text{m}$ | \pm | 5.8E-01 μm |
| 523.0 μm | 522.0 μm | -1.0 μm | $\pm 5 \mu\text{m}$ | \pm | 5.8E-01 μm |