

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.10017	-0.170	± 0.2	±	4.3E-01 mm						
0 a 0.2	0.20018	-0.180	± 0.2	±	4.3E-01 mm						
0 a 0.3	0.30018	-0.180	± 0.2	±	4.3E-01 mm						
0 a 0.4	0.40013	-0.130	± 0.2	±	4.3E-01 mm						
0 a 0.5	0.50014	-0.140	± 0.3	±	4.3E-01 mm						
0 a 0.6	0.60013	-0.130	± 0.3	±	4.3E-01 mm						
0 a 0.7	0.70014	-0.140	± 0.3	±	4.3E-01 mm						
0 a 0.8	0.80003	-0.030	± 0.3	±	4.3E-01 mm						
0 a 0.9	0.90002	-0.020	± 0.3	±	4.3E-01 mm						
0 a 1	0.99992	0.080	± 0.4	±	4.3E-01 mm						