

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.161 °C	-0.3 °C	-0.139 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.025 °C	19.8 °C	-0.225 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.899 °C	40.1 °C	0.201 °C	± 0.5 °C	± 4.0E-01 °C