

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
300.6 RPM	300 RPM	-0.6 RPM	- ±	2.9E+02 RPM
801.2 RPM	800 RPM	-1.2 RPM	- ±	2.9E+02 RPM
1302 RPM	1300 RPM	-2 RPM	- ±	2.9E+02 RPM
1803 RPM	1800 RPM	-3 RPM	- ±	2.9E+02 RPM
2305 RPM	2300 RPM	-5 RPM	- ±	2.9E+02 RPM