

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
300.1 RPM	300 RPM	-0.1 RPM	-	±	2.9E+02 RPM
799.5 RPM	800 RPM	0.5 RPM	-	±	2.9E+02 RPM
1301 RPM	1300 RPM	-1 RPM	-	±	2.9E+02 RPM
1802 RPM	1800 RPM	-2 RPM	-	±	2.9E+02 RPM
2303 RPM	2300 RPM	-3 RPM	-	±	2.9E+02 RPM