

Certificado No. (Certificate #):		53298		Equipo (Instrument): Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-01-22		Marca (Brand): PEL			
Patrón (Standard)		UBP (UUT)		Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
25.0000	mm	25.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
27.0000	mm	27.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
28.0000	mm	28.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
30.0000	mm	30.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
35.0000	mm	35.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
40.0000	mm	40.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
42.0000	mm	42.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
44.0000	mm	44.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
45.0000	mm	45.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
48.0000	mm	48.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm
50.0000	mm	50.0000	mm	0.0000 μm	± 4 μm	±	5.8E+00 μm