

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
584.5 RPM	600 RPM	15.5 RPM	± 2 RPM	±	8.2E-02 RPM
1260.6 RPM	1300 RPM	39.4 RPM	± 2 RPM	±	1.1E-01 RPM