

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.190 °C	-24.5 °C	0.690 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.611 °C	1.0 °C	0.389 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.415 °C	5.7 °C	0.285 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.589 °C	10.6 °C	0.011 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.333 °C	30.2 °C	-0.133 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.030 °C	40.1 °C	0.070 °C	± 0.45 °C	± 4.0E-01 °C

