

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.190 °C	-25.2 °C	-0.010 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	0.611 °C	0.7 °C	0.089 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.415 °C	5.5 °C	0.085 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.589 °C	10.5 °C	-0.089 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.5 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.333 °C	30.4 °C	0.067 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.030 °C	40.4 °C	0.370 °C	± 0.5 °C	± 4.0E-01 °C

