

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.190 °C	-24.8 °C	0.390 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	0.611 °C	0.8 °C	0.189 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.415 °C	5.4 °C	-0.015 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.589 °C	10.5 °C	-0.089 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.333 °C	30.1 °C	-0.233 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.030 °C	39.9 °C	-0.130 °C	± 0.5 °C	± 4.0E-01 °C

