

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.982 °C	30.0 °C	0.018 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.323 %HR	30.1 %HR	0.777 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	50.1 %HR	0.573 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	71.255 %HR	71.9 %HR	0.645 %HR	± 2.0 %HR	± 1.9E+00 %HR