

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0625 in	0.062 in	-0.0005 in	± 0.002 in	±	5.8E-04 in
0.100 in	0.100 in	0.0000 in	± 0.002 in	±	5.8E-04 in
0.125 in	0.125 in	0.0000 in	± 0.002 in	±	5.8E-04 in
0.200 in	0.200 in	0.0000 in	± 0.002 in	±	5.8E-04 in
0.250 in	0.250 in	0.0000 in	± 0.002 in	±	5.8E-04 in
0.300 in	0.300 in	0.0000 in	± 0.002 in	±	5.8E-04 in
0.500 in	0.500 in	0.0000 in	± 0.002 in	±	5.8E-04 in
1 in	1.000 in	0.0000 in	± 0.002 in	±	5.8E-04 in
2 in	2.000 in	0.0000 in	± 0.002 in	±	5.8E-04 in