

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.1	0.09987	0.130	± 0.6	±	4.3E-01 mm						
0 a 0.2	0.19984	0.160	± 0.6	±	4.3E-01 mm						
0 a 0.3	0.29979	0.210	± 0.7	±	4.3E-01 mm						
0 a 0.4	0.39972	0.280	± 0.7	±	4.3E-01 mm						
0 a 0.5	0.49968	0.320	± 0.8	±	4.3E-01 mm						
0 a 0.6	0.59960	0.400	± 0.8	±	4.3E-01 mm						
0 a 0.7	0.69951	0.490	± 0.8	±	4.3E-01 mm						
0 a 0.8	0.79952	0.480	± 0.9	±	4.3E-01 mm						
0 a 0.9	0.89949	0.510	± 0.9	±	4.3E-01 mm						
0 a 1	0.99939	0.610	± 1	±	4.3E-01 mm						