

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99904	0.960	± 1	±	4.3E-01 mm						
0 a 2	1.99914	0.860	± 1.4	±	4.8E-01 mm						
0 a 3	2.99923	0.770	± 1.8	±	4.8E-01 mm						
0 a 4	3.99927	0.730	± 2.2	±	4.8E-01 mm						
0 a 5	4.99915	0.850	± 2.6	±	4.8E-01 mm						
0 a 6	5.99926	0.740	± 3	±	3.6E-01 mm						
0 a 7	6.99926	0.740	± 3.4	±	3.6E-01 mm						
0 a 8	7.99936	0.640	± 3.8	±	3.6E-01 mm						
0 a 9	8.99947	0.530	± 4.2	±	3.6E-01 mm						
0 a 10	9.99986	0.140	± 4.6	±	3.6E-01 mm						