

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.022 °C	30.1 °C	0.078 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	33.617 %HR	35.5 %HR	1.883 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.828 %HR	53.1 %HR	1.272 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.828 %HR	71.6 %HR	-0.228 %HR	± 3.0 %HR	± 1.9E+00 %HR