

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
23.3 μm	23 μm	-0.30 μm	± 3 μm	±	5.8E-01 μm
51.6 μm	52 μm	0.40 μm	± 3 μm	±	5.8E-01 μm
124.6 μm	125 μm	0.40 μm	± 3 μm	±	6.1E-01 μm
250.3 μm	251 μm	0.70 μm	± 3 μm	±	6.9E-01 μm
486.0 μm	486 μm	0.00 μm	± 3 μm	±	9.4E-01 μm
963.0 μm	963 μm	0.00 μm	± 3 μm	±	1.6E+00 μm
1,000.0 μm	1,000 μm	0.00 μm	± 3 μm	±	1.6E+00 μm