

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.022 °C	30.1 °C	0.078 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	33.617 %HR	35.1 %HR	1.483 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.828 %HR	51.7 %HR	-0.128 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.828 %HR	70.2 %HR	-1.628 %HR	± 3.0 %HR	± 1.9E+00 %HR