

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 °C	28.297 °C	28.6 °C	0.303 °C	± 1.0 °C	± 4.0E-01 °C
25 %HR	24.282 %HR	29 %HR	4.718 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	44.834 %HR	49 %HR	4.166 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.529 %HR	80 %HR	4.471 %HR	± 5.0 %HR	± 1.9E+00 %HR