

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
28 °C	28.297 °C	28.0 °C	-0.297 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	24.282 %HR	21 %HR	-3.282 %HR	± 10.0 %HR	± 1.9E+00 %HR
45 %HR	44.834 %HR	36 %HR	-8.834 %HR	± 10.0 %HR	± 1.9E+00 %HR
75 %HR	75.529 %HR	66 %HR	-9.529 %HR	± 10.0 %HR	± 1.9E+00 %HR