

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
28 °C	28.117 °C	27.8 °C	-0.317 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	24.138 %HR	22 %HR	-2.138 %HR	± 10.0 %HR	± 1.9E+00 %HR
45 %HR	44.732 %HR	36 %HR	-8.732 %HR	± 10.0 %HR	± 1.9E+00 %HR
75 %HR	75.519 %HR	71 %HR	-4.519 %HR	± 10.0 %HR	± 1.9E+00 %HR