

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 °C	28.538 °C	28.5 °C	-0.038 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	24.271 %HR	20 %HR	-4.271 %HR	± 10.0 %HR	± 1.9E+00 %HR
45 %HR	44.353 %HR	36 %HR	-8.353 %HR	± 10.0 %HR	± 1.9E+00 %HR
75 %HR	75.069 %HR	63 %HR	-12.069 %HR	± 10.0 %HR	± 1.9E+00 %HR