

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 °C	27.806 °C	26 °C	-1.806 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	23.392 %HR	24 %HR	0.608 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	44.629 %HR	43 %HR	-1.629 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.529 %HR	72 %HR	-3.529 %HR	± 5.0 %HR	± 1.9E+00 %HR