

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
28 °C	28.538 °C	30 °C	1.462 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	24.271 %HR	25 %HR	0.729 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	44.353 %HR	44 %HR	-0.353 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.069 %HR	72 %HR	-3.069 %HR	± 5.0 %HR	± 1.9E+00 %HR