

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
28 °C	28.538 °C	28 °C	-0.538 °C	± 2.0 °C	± 4.0E-01 °C
25 %HR	24.271 %HR	22 %HR	-2.271 %HR	± 5.0 %HR	± 1.9E+00 %HR
45 %HR	44.353 %HR	42 %HR	-2.353 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.069 %HR	72 %HR	-3.069 %HR	± 5.0 %HR	± 1.9E+00 %HR