

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
2 °C	2.286 °C	1.9 °C	-0.386 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.013 °C	5.4 °C	0.387 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.082 °C	8.0 °C	-0.082 °C	± 0.5 °C	± 4.0E-01 °C