

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.286 °C	2.2 °C	-0.086 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.013 °C	5.1 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.082 °C	8.0 °C	-0.082 °C	± 0.5 °C	± 4.0E-01 °C