

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.286 °C	2.2 °C	-0.086 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.013 °C	4.7 °C	-0.313 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.082 °C	8.2 °C	0.118 °C	± 0.5 °C	± 4.0E-01 °C