

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.040 °C	22.1 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.7 °C	0.250 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.085 °C	31.0 °C	-0.085 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	47.5 %HR	-0.831 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.769 %HR	59.0 %HR	1.231 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	69.0 %HR	1.620 %HR	± 3.0 %HR	± 1.9E+00 %HR