

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.040 °C	22.0 °C	-0.040 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.5 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.085 °C	31.2 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	48.5 %HR	0.169 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.769 %HR	58.4 %HR	0.631 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	67.6 %HR	0.220 %HR	± 3.0 %HR	± 1.9E+00 %HR