

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.040 °C	22.1 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.6 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.085 °C	31.0 °C	-0.085 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	49.0 %HR	0.669 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.769 %HR	58.3 %HR	0.531 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	67.7 %HR	0.320 %HR	± 3.0 %HR	± 1.9E+00 %HR