

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 22 °C                      | 22.040 °C                  | 21.7 °C            | -0.340 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.450 °C                  | 25.3 °C            | -0.150 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 31 °C                      | 31.085 °C                  | 31.0 °C            | -0.085 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 48 %HR                     | 48.331 %HR                 | 48.3 %HR           | -0.031 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 57 %HR                     | 57.769 %HR                 | 58.0 %HR           | 0.231 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 67 %HR                     | 67.380 %HR                 | 67.8 %HR           | 0.420 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |