

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
22 °C	22.040 °C	22.1 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.5 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
31 °C	31.085 °C	31.1 °C	0.015 °C	± 0.5 °C	± 4.0E-01 °C
48 %HR	48.331 %HR	49.2 %HR	0.869 %HR	± 3.0 %HR	± 1.9E+00 %HR
57 %HR	57.769 %HR	58.7 %HR	0.931 %HR	± 3.0 %HR	± 1.9E+00 %HR
67 %HR	67.380 %HR	67.0 %HR	-0.380 %HR	± 3.0 %HR	± 1.9E+00 %HR