

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 22 °C                      | 22.040 °C                  | 22.0 °C            | -0.040 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.450 °C                  | 25.5 °C            | 0.050 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 31 °C                      | 31.085 °C                  | 31.2 °C            | 0.115 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 48 %HR                     | 48.331 %HR                 | 48.9 %HR           | 0.569 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 57 %HR                     | 57.769 %HR                 | 57.1 %HR           | -0.669 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 67 %HR                     | 67.380 %HR                 | 66.9 %HR           | -0.480 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |